
CERTIFICATE OF ANALYSIS

Product:	Article number(s):
WorkBeads™ 100S	10 200 001
	10 200 002
	10 200 005
Lot number: 153046	10 200 010
	10 200 050
	10 200 060
Storage solution:	20% ethanol, 0.2 M sodium acetate

	Measured Value	Specification
Ionic capacity:	205 µmol H ⁺ /ml	180-250 µmol H ⁺ /ml
Average particle size, Dv50:	99 µm	90-110 µm
Flow stability: (linear flow rate in a 25 x 200 mm column)*	≥1000 cm/hr	≥1000 cm/hr

*Property measured on base matrix.

Release date (year-month-day): 2023-11-15	Expiry date (year-month): 2026-11
Retest date (year-month-day): 2023-11-14	Manufacturing date (year-month): 2018-11

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2023-11-15

By Åsa Nilsson

DocuSigned by:



FD9074A2ACBC431...

Quality Assurance
