

## CERTIFICATE OF ANALYSIS

---

Product: **WorkBeads™ 40 Q**

Article number: **40 100 001, 40 100 002, 40 100 010, 40 100 050,  
40 100 060, 40 100 210, 40 100 250**

Lot number: **153049**

Product release date: **2018 NOV 16**

Expire date: **2023-11**

---

	<b>Measured values</b>	<b>Specifications</b>
Ionic capacity	203 µmol Cl <sup>-</sup> /ml	180-250 µmol Cl <sup>-</sup> /ml
Average particle size (Dv50)	57 µm	32-60 µm
Flow properties (linear flow in a 10x300 mm column)	>10 cm/min	≥10 cm/min

---

Issued 2019, SEP 20

by

  
Lars Haneskog  
Director of Quality Control