

## CERTIFICATE OF ANALYSIS

---

Product: **WorkBeads™ 200 SEC SG**

Article number: **20 300 210, 20 300 250**

Lot number: **153066**

Manufacturing date: **2018 DEC 03**

---

	Measured values	Specifications
Average particle size, Dv50:	170 µm	150 -190 µm
Volume fraction of particles within the size range of 130-250 µm:	85.6 %	≥ 85 %
Kd (Ovalbumin):	0.69	0.62 - 0.75

---

Issued 2018 DEC 05

by



---

Mats Johnson  
Chief Executive officer