

CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40S**

Article number: **40 200 001, 40 200 002, 40 200 005, 40 200 010,
40 200 050, 40 200 060, 40 200 210, 40 200 250**

Lot number: **153084**

Product release date: **2018 DEC 10**

Expire date: **2023-12**

	Measured values	Specifications
Ionic capacity	197 µmol H ⁺ /mL	180 – 250 µmol H ⁺ /mL
Average particle size (Dv50)	53.6 µm	32 - 60 µm
Flow properties (linear flow in a 10x300 mm column)	>10 cm/min	≥10 cm/min

Issued **2020 FEB 04**

by



Helena Pahlm
Quality Assurance Manager