

CERTIFICATE OF ANALYSIS

Product:	BabyBio ACT 1 ml
Article number:	45 400 001, 45 400 002, 45 400 003, 45 400 004
Lot number:	153140
Resin lot number:	153154
Product release date:	2019 MAR 29
Expire date:	2022-03

	Measured values	Specifications
Ligand Conc. of reactive groups:	273 $\mu\text{mol/ml}$	> 200 $\mu\text{mol/ml}$
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued 2019 MAR 29
by



Mats Johnson
CEO