

## CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40/1000 SEC**

Article number: **40 300 001, 40 300 003, 40 300 010, 40 300 050,  
40 300 60, 40 300 201, 40 300 203, 40 300 250,  
40 300 260**

Lot number: **153146**

Manufacturing date: **2019 MAR 05**

	Measured values	Specifications
Average particle size (Dv50)	54.2 µm	32 - 60 µm
Kd (Ovalbumin)	0.65	0.62 - 0.75
Flow properties (linear flow in a 10x300 mm column)	>10 cm/min	≥10 cm/min

Issued 2019 MAR 13

by



**Mats Johnson**  
Chief Executive officer