
CERTIFICATE OF ANALYSIS

Product: **BabyBio ACT 5 ml**

Article number: **45 400 005, 45 400 006, 45 400 007, 45 400 008**

Lot number: **153155**

Resin lot number: **153154**

Product release date: **2019 MAY 13**

Expire date: **2022-03**

	Measured values	Specifications
Conc. of reactive groups:	273 µmol/ml	> 200 µmol/ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued 2019 MAY 13

by



Mats Johnson
Chief Executive officer