
CERTIFICATE OF ANALYSIS

Product:	BabyBio™ A 1 ml
Article number:	45 605 101, 45 605 102, 45 605 103, 45 605 104
Lot number:	153156
Resin lot number:	153249
Product release date:	2019 MAY 09
Expire date:	2022-05

	Measured values	Specifications
Dynamic binding capacity for IgG ¹ .	50.8 mg/ml	> 40 mg/ml
Gap control: (4 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

1. Test done at 2.5 minutes residence time in 6.6x100 mm column.

Issued 2019 MAY 13

by



Mats Johnson
Chief Executive officer