

CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ Protein A**

Article number: **40 605 001, 40 605 002, 40 605 003,
40 605 004, 40 605 005**

Lot number: **153249**

Product release date: **2019 MAY 07**

Expire date: **2022-07**

	Measured values	Specifications
Dynamic binding capacity (at 2.5 minutes residence time)	50.8 mg IgG/ml resin	>40 mg IgG/ml resin
Flow Stability (Linear flow in a 10 x 300 mm column)	300 cm/h	≥300 cm/h
Average particle size, Dv50	46.9 µm	32-60 µm
Particle size range within 30-75 µm	85 %	≥ 80%

Issued 2019 MAY 13

by



Mats Johnson
Chief Executive officer