

## CERTIFICATE OF ANALYSIS

---

Product: **WorkBeads™ 40/100 SEC**

Article number: **40 340 001, 40 340 003, 40 340 010, 40 340 050,  
40 340 210, 40 340 250, 40 340 260**

Lot number: **153382**

Product release date: **2019 JUL 04**

Expire date: **2022-07**

---

	Measured values	Specifications
Average particle size (Dv50):	56.4 µm	32 - 60 µm
Kd (Ovalbumin):	0.56	0.40 - 0.60
Flow properties (linear flow in a 10x300 mm column):	>10 cm/min	≥10 cm/min

---

Issued 2019 JUL 18

by



Lars Haneskog  
Director of Quality Control