


CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40 SEC Cube**
Article number: **40 301 050**
Lot number: **153389**
Product release date: **2019-09-20**
Expire date: **2024-09**

	Measured values	Specifications
Average particle size (Dv50)	56 µm	32 - 60 µm
K _D (Thyroglobulin)	0.26	0.20 – 0.30
Flow properties (linear flow in a 10 ×300 mm column)	>10 cm/min	≥10 cm/min

Issued Oct. 15, 2019

by



Lars Haneskog
Director Quality control