
CERTIFICATE OF ANALYSIS

Product: **BabyBio™ A 5 ml**

Article number: **45 605 105, 45 605 106, 45 605 107, 45 605 108**

Lot number: **153393**

Resin lot number: **153249**

Product release date: **2019 JUL 03**


Expire date: **2022-05**

	Measured values	Specifications
Dynamic binding capacity for IgG ¹ .	50.8 mg/ml	> 40 mg/ml
Gap control: (15 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (17 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

1. Test done at 2.5 minutes residence time in 6.6x100 mm column.

Issued 2019 JUL 03

by



Lars Haneskog
Director Quality control