

CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40/10 000 SEC**

Article number: **40 350 001, 40 350 003, 40 350 010, 40 350 050,
40 350 203, 40 350 210, 40 350 250, 40 350 260**

Lot number: **153394**


Product release date: **2019 JUL 04**

Expire date: **2024-07**

	Measured values	Specifications
Average particle size (Dv50):	50 µm	32 - 60 µm
Kd (Ovalbumin):	0.76	0.65 - 0.85
Flow properties (linear flow in a 10x300 mm column):	>10 cm/min	≥10 cm/min

Issued 2019 JULY 04

by



Lars Haneskog
Director Quality control