

CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40 DEAE**

Article number: **40 150 001, 40 150 002, 40 150 010, 40 150 050,
40 150 210, 40 150 250**

Lot number: **153416**


Product release date: **2019 JUL 10**

Expire date: **2024-07**

	Measured values	Specifications
Ionic capacity:	153 µmol Cl ⁻ /ml	110 - 160 µmol Cl ⁻ /ml
Average particle size (Dv50):	50.6 µm	32 - 60 µm
Flow properties (linear flow in a 10x300 mm column):	>10 cm/min	≥10 cm/min

Issued 2019 JUL 11

by



Lars Haneskog
Director of Quality Control