

## CERTIFICATE OF ANALYSIS

---

Product: **WorkBeads™ 40 DEAE**

Article number: **40 150 001, 40 150 002, 40 150 010, 40 150 050,  
40 150 210, 40 150 250**

Lot number: **BH 021 (153419)**

Product release date: **2019 JUL 15**

Expire date: **2023-07**

---

	<b>Measured values</b>	<b>Specifications</b>
Ionic capacity:	139 µmol Cl <sup>-</sup> /ml	110 - 160 µmol Cl <sup>-</sup> /ml
Average particle size (Dv50):	49 µm	32 - 60 µm
Flow properties (linear flow in a 10x300 mm column):	> 10 cm/min	≥10 cm/min

---

Issued 2019 JUL 15

by



Lars Haneskog  
Director of Quality Control