

CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40/1000 SEC**

Article number: **40 300 001, 40 300 003, 40 300 010, 40 300 050,
40 300 60, 40 300 201, 40 300 203, 40 300 250,
40 300 260**

Lot number: **153444**

Product release date: **2019 SEP 20**

Expire date: **2024-10**

	Measured values	Specifications
Average particle size (Dv50)	55 µm	32 - 60 µm
Kd (Ovalbumin)	0.66	0.62 - 0.75
Flow properties (linear flow in a 10x300 mm column)	>10 cm/min	≥10 cm/min

Issued 2019 SEP 20

by



Lars Haneskog
Director of Quality Control