

---

## CERTIFICATE OF ANALYSIS

---

**Product:** BabyBio TREN 1 ml

Article number: 45 655 211, 45 655 212, 45 655 213, 45 655 214

Lot number: 153457

Resin lot number: 151845

Product release date: 2016 SEP 14

Expire date: 2021-09


---

	Measured values	Specifications
Metal ion capacity (Tested with Cu <sup>2+</sup> )	59 µmol Cu <sup>2+</sup> /ml	50 – 60 µmol Cu <sup>2+</sup> /ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

---

Issued **2019 SEP 30**

by



Lars Haneskog  
Director of Quality Control