

## CERTIFICATE OF ANALYSIS

---

Product: **WorkBeads™ 40 DEAE**

Article number: **40 150 001, 40 150 002, 40 150 010, 40 150 050,  
40 150 210, 40 150 250**

Lot number: **153469**

Product release date: **2019 SEP 20**


Expire date: **2024-09**

---

	<b>Measured values</b>	<b>Specifications</b>
Ionic capacity:	136 µmol Cl <sup>-</sup> /ml	110 - 160 µmol Cl <sup>-</sup> /ml
Average particle size (Dv50):	54 µm	32 - 60 µm
Flow properties (linear flow in a 10x300 mm column):	>10 cm/min	≥10 cm/min

---

Issued 2019 SEP 20  
by



Lars Haneskog  
Director of Quality Control