

CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40S**

Article number: **40 200 001, 40 200 002, 40 200 005, 40 200 010,
40 200 050, 40 200 060, 40 200 210, 40 200 250**

Lot number: **153470**

Product release date: **2019 AUG 23**

Expire date: **2024-08**

| | Measured values | Specifications |
|--|-----------------------------|-----------------------------------|
| Ionic capacity | 202 µmol H ⁺ /mL | 180 – 250 µmol H ⁺ /mL |
| Average particle size (Dv50) | 55 µm | 32 - 60 µm |
| Flow properties (linear flow in a 10x300 mm column) | >10 cm/min | ≥10 cm/min |

Issued **2019 SEP 23**

by



Lars Haneskog
Director of Quality Control