

CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40S**

Article number: **40 200 001, 40 200 002, 40 200 005, 40 200 010,
40 200 050, 40 200 060, 40 200 210, 40 200 250**

Lot number: **153486**


Product release date: **2019 SEP 06**

Expire date: **2024-06**

	Measured values	Specifications
Ionic capacity	195 $\mu\text{mol H}^+/\text{mL}$	180 – 250 $\mu\text{mol H}^+/\text{mL}$
Average particle size (Dv50)	53 μm	32 - 60 μm
Flow properties (linear flow in a 10x300 mm column)	>10 cm/min	≥ 10 cm/min

Issued **2019 SEP 23**

by


Lars Haneskog
Director of Quality Control