

## CERTIFICATE OF ANALYSIS

---

Product: **WorkBeads™ Macro SEC**  
Article number: **CP40370001, CP40370003, CP40370010,  
CP40370050**  
Lot number: **153501**  
Product release date: **2019 NOV 1**  
Expire date: **2024-10**

---

	Measured values	Specifications
Average particle size (Dv50)	54 µm	32 - 60 µm
K <sub>D</sub> (Thyroglobulin)	0.50	0.39 – 0.51
Flow properties (linear flow in a 10x300 mm column)	11.5 cm/min	≥10 cm/min

---

Issued 2020 JAN 13

by



Lars Haneskog  
Director Quality Control