

CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40 Q**

Article number: **40 100 001, 40 100 002, 40 100 010, 40 100 050,
40 100 060, 40 100 210, 40 100 250**

Lot number: **153700**

Product release date: **2019 DEC 05**

Expire date: **2024-12**

| | Measured values | Specifications |
|--|------------------------------|----------------------------------|
| Ionic capacity | 205 µmol Cl ⁻ /ml | 180-250 µmol Cl ⁻ /ml |
| Average particle size (Dv50) | 59 µm | 32-60 µm |
| Flow properties (linear flow in a 10x300 mm column) | >10 cm/min | ≥10 cm/min |

Issued **2019 DEC 10**

by



Lars Haneskog
Director of Quality Control