

## CERTIFICATE OF ANALYSIS

---

Product: **WorkBeads™ Protein A**

Article number: **40 605 001, 40 605 002, 40 605 003,  
40 605 004, 40 605 005**

Lot number: **153702**

Product release date: **2019 DEC 12**


Expire date: **2022-12**

---

	<b>Measured values</b>	<b>Specifications</b>
Dynamic binding capacity (at 2.5 minutes residence time)	54 mg IgG/ml resin	> 40 mg IgG/ml resin
Flow Stability (Linear flow in a 10 x 300 mm column)	300 cm/h	≥300 cm/h
Average particle size, Dv50	45 µm	32-60 µm
Particle size range within 30-75 µm	90 %	≥ 80%

---

Issued 2019 DEC 17  
by



Helena Pahlm  
Quality Assurance Manager