

## CERTIFICATE OF ANALYSIS

---

Product: **WorkBeads™ 40/10 000 ACT**

Article number: **40 450 001, 40 450 003, 40 450 010, 40 450 050,  
40 450 210, 40 450 250**

Lot number: **153703**

Product release date: **2019 DEC 10**

Expire date: **2022-12**

---

	<b>Measured values</b>	<b>Specifications</b>
Average particle size (Dv50):	49 µm	32 - 60 µm
Conc. of reactive groups:	209 µmol/ml	> 200 µmol/ml
Flow properties (linear flow in a 10x300 mm column):	>10 cm/min	≥ 10 cm/min

---

Issued **2019 DEC 10**

by



Lars Haneskog  
Director of Quality Control