

CERTIFICATE OF ANALYSIS

Product:	WorkBeads 40Q HS 50/20
Article number:	CU55410501
Column ID:	153714
Resin lot number:	153700
Expiry:	2024-12

	Measured values
Efficiency ¹	7100 N/m
Asymmetry ¹	1.3
Back pressure over column at recommended maximum flow rate ²	0.13 MPa
Bed height	20.1 cm


1. Tested at 30 cm/h (9.8 ml/min) with 4000 µl 2% acetone.

2. Tested at 275 cm/h (90 ml/min) with water at 25°C.

Note: The column is delivered in 20% ethanol. In order to prevent too high back pressure, avoid flow rates above 45 ml/min before the storage solution has been replaced.

Issued December 12, 2019

by



Lars Haneskog

Director Quality Control