

---

## CERTIFICATE OF ANALYSIS

---

Product: **WorkBeads™ affimAb**

Article number: **40 800 001, 40 800 002, 40 800 010,  
40 800 050, 40 200 060**

Lot number: **153721**

Product release date: **2019 DEC 18**

Expire date: **2022-12**

---

	<b>Measured values</b>	<b>Specifications</b>
Dynamic binding capacity: (at 2.5 minutes residence time)	43 mg IgG/ml resin	>40 mg IgG/ml resin
Flow Stability: (Linear flow in a 10 x 300 mm column)	>10 cm/min	≥10 cm/min
Average particle size, Dv50:	53 µm	32-60 µm
Particle size range within 30-75 µm:	92 %	≥ 80%

---

Issued 2019 DEC 18

by



Lars Haneskog  
Director of Quality Control