

CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40 DEAE**

Article number: **40 150 001, 40 150 002, 40 150 010, 40 150 050,
40 150 210, 40 150 250**

Lot number: **153744**

Product release date: **2020 JAN 27**

Expire date: **2025-01**

	Measured values	Specifications
Ionic capacity:	128 µmol Cl ⁻ /ml	110 - 160 µmol Cl ⁻ /ml
Average particle size (Dv50):	53 µm	32 - 60 µm
Flow properties (linear flow in a 10x300 mm column):	>10 cm/min	≥10 cm/min

Issued 2020 JAN 28
by



Helena Pahlm
Quality Assurance Manager