

## CERTIFICATE OF ANALYSIS

---

Product: **WorkBeads™ 40 IDA**

Article number: **40 601 001, 40 601 003, 40 601 010, 40 601 050**

Lot number: **153763**

Product release date: **2020 FEB 06**

Expire date: **2025-02**

---

	<b>Measured values</b>	<b>Specifications</b>
Capacity:	65 µmol Cu <sup>2+</sup> /mL	50 - 70 µmol Cu <sup>2+</sup> /mL
Average particle size (Dv50):	56 µm	32 - 60 µm
Flow properties (linear flow in a 10x300 mm column):	≥10 cm/min	≥10 cm/min

---

Issued 2020 FEB 06  
by



Helena Panlm  
Quality Assurance Manager