

## CERTIFICATE OF ANALYSIS

**Product:** BabyBio IDA 1 ml

Article number: 45 655 011, 45 655 012, 45 655 013, 45 655 014

Lot number: 153767      obsolete due to name change: CC-22-10384

Resin lot number: 153763      Philip Eriksson  
2022-09-30

Product release date: 2020 FEB 06

Expire date: 2025-02

DocuSigned by:  
*Philip Eriksson*  
158C686C7E5E4A0...

	Measured values	Specifications
Metal ion capacity (Tested with Cu <sup>2+</sup> )	65 µmol Cu <sup>2+</sup> /ml	50 – 70 µmol Cu <sup>2+</sup> /ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued 2020 FEB 11  
by



\_\_\_\_\_  
Helena Pahlm  
Quality Assurance Manager