

---

## CERTIFICATE OF ANALYSIS

---

<b>Product:</b>	<b>BabyBio IDA 1 ml</b>
Article number:	<b>45 655 011, 45 655 012, 45 655 013, 45 655 014</b>
Lot number:	<b>153767</b>
Resin lot number:	<b>153763</b>
Product release date:	<b>2020 FEB 06</b>
Expire date:	<b>2025-02</b>

---

	<b>Measured values</b>	<b>Specifications</b>
Metal ion capacity (Tested with Cu <sup>2+</sup> )	65 µmol Cu <sup>2+</sup> /ml	50 – 70 µmol Cu <sup>2+</sup> /ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

---

Issued 2020 FEB 11  
by



---

Helena Pahlm  
Quality Assurance Manager