
CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini IDA 1 mL	Article number(s):	45 655 011 45 655 012 45 655 013 45 655 014
Lot number:	153767		
Storage solution:			20% ethanol

	Measured Value	Specification
Metal binding capacity*:	65 µmol Cu ²⁺ /mL	50-70 µmol Cu ²⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2020-02-06

Expiry date (year-month): 2025-02

Manufacturing date (year-month): 2020-02

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-09-30

By Philip Eriksson

Quality Assurance
