

## CERTIFICATE OF ANALYSIS

**Product:** **BabyBio S 1 ml**

**Article number:** **45 200 101, 45 200 102, 45 200 103, 45 200 104**

**Lot number:** **153778**      obsolete due to name change CC-22-10384

**Resin lot number:** **153470**      Philip Eriksson  
2022-08-31


**Product release date:** **2019 AUG 23**

**Expire date:** **2024-08**

DocuSigned by:  
*Philip Eriksson*  
158C686C7E5E4A0...

	<b>Measured values</b>	<b>Specifications</b>
Ionic capacity	202 $\mu\text{mol Na}^+/\text{ml}$	180-250 $\mu\text{mol Na}^+/\text{ml}$
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued 2020 FEB 28  
by



Helena Pahlm  
Quality Assurance Manager