

CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ Protein A**

Article number: **40 605 001, 40 605 002, 40 605 003,
40 605 004, 40 605 005**

Lot number: **153782**


Product release date: **2020 APR 27**

Expire date: **2023-04**

	Measured values	Specifications
Dynamic binding capacity (at 2.5 minutes residence time)	56 mg IgG/ml resin	> 40 mg IgG/ml resin
Flow Stability (Linear flow in a 10 x 300 mm column)	>300 cm/h	≥300 cm/h
Average particle size, Dv50	47 µm	32-60 µm
Particle size range within 30-75 µm	87 %	≥ 80%

Issued **2020 APR 27**

by



Johanna Tawe
QMS Director