

## CERTIFICATE OF ANALYSIS

**Product:** BabyBio IDA 5 ml

Article number: 45 655 015, 45 655 016, 45 655 017, 45 655 018

Lot number: 153827

Resin lot number: 153763


Product release date: 2020 FEB 06

Expire date: 2025-02

	Measured values	Specifications
Metal ion capacity (Tested with Cu <sup>2+</sup> )	65 µmol Cu <sup>2+</sup> /ml	50 – 70 µmol Cu <sup>2+</sup> /ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued **2020 JUN 22**

by



Helena Pahlm  
Quality Assurance Manager