
CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 100Q**

Article number: **10 100 001, 10 100 002, 10 100 005, 10 100 010,
10 100 050, 10 100 060**

Lot number: **153854**

Product release date: **2020 JUL 02**

Expire date: **2021-11**

	Measured values	Specifications
Ionic capacity:	175 µmol Cl ⁻ /ml	140 - 200 µmol Cl ⁻ /ml
Average particle size (Dv50):	100 µm	90 -110 µm
Flow acceptance (in a 25x200 mm column):	>1000 cm/h	≥1000 cm/h

Issued **2020 JUL 02**

by



Helena Pahlm
Quality Assurance Manager