

## CERTIFICATE OF ANALYSIS

|                       |   |
|-----------------------|---|
| <b>Product:</b>       | <b>BabyBio™ affimAb 5 ml</b>                          |
| Article number:       | <b>45 800 105, 45 800 106, 45 800 107, 45 800 108</b> |
| Lot number:           | <b>153880</b>   |
| Resin lot number:     | <b>153721</b>   |
| Product release date: | <b>2019 DEC 18</b>                                    |
| Expire date:          | <b>2022-12</b>  |

|   | <b>Measured values</b> | <b>Specifications</b> |
|---|------------------------|-----------------------|
| Dynamic binding capacity for IgG <sup>1</sup> . | 43 mg/ml               | > 40 mg/ml            |
| Gap control:<br>(15 ml/min, water, 22 °C)       | Passed                 | No gap                |
| Maximum flow test:<br>(17 ml/min, water, 22 °C) | Passed                 | Stable pressure       |
| Visual control:                                 | Passed                 | No visible defects    |

1. Test done at 2.5 minutes residence time in 6.6x100 mm column.

Issued **2020 MAY 07**

by



Johanna Tawe  
QMS Director