
CERTIFICATE OF ANALYSIS

Product:	BabyBio S 5 ml
Article number:	45 200 105, 45 200 106, 45 200 107, 45 200 108
Lot number:	153898
Resin lot number:	153470
Product release date:	2019 AUG 23
Expire date:	2024-08

	Measured values	Specifications
Ionic capacity:	202 $\mu\text{mol Na}^+/\text{ml}$	180-250 $\mu\text{mol Na}^+/\text{ml}$
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued **2020 MAY 13**

by



Johanna Tawe
QMS Director