

CERTIFICATE OF ANALYSIS

Product:	BabyBio NiMAC 1 ml
Storage solution:	20% ethanol
Article number:	45 655 311, 45 655 312, 45 655 313, 45 655 314
Lot number:	153909
Resin lot number:	80001-3029-069
Manufacturing date:	2020 APR
Product release date:	2020 APR 30
Expire date:	2025-04

	Measured values	Specifications
Cu ²⁺ binding capacity before charging with Ni ²⁺	125 µmol/ml	> 60 µmol/ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2020 JUN 30**

by



Helena Pahlm
Quality Assurance Manager