
CERTIFICATE OF ANALYSIS

Product: **OptioBio™ 40Q 10x100**

Article number: **55 410 011**

Column ID: **153943-04**

Resin lot number: **153921**

Product release date: **2019 SEP 20**

Expire date: **2024-09**

	Measured values	Specifications
Efficiency	8794 N/m	> 6350 N/m
Asymmetry	1.18	0.75 - 1.50
Column backpressure at max recommended flow (450 cm/h, 6 ml/min, water)	0.06 MPa	Not specified
Bed height	101 mm	100 - 102 mm

Issued **2020 JUN 16**

by



Helena Palm
Quality Assurance Manager