

---

## CERTIFICATE OF ANALYSIS

---

**Product:** BabyBio Q 5 ml

Article number: 45 100 105, 45 100 106, 45 100 107, 45 100 108

Lot number: 153989

Resin lot number: 153921

Product release date: 2020 MAY 15

Expire date: 2023-11

---

	Measured values	Specifications
Ionic capacity:	203 $\mu\text{mol Cl}^-/\text{ml}$	180-250 $\mu\text{mol Cl}^-/\text{ml}$
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

---

Issued **2020 JUL 06**

by



---

Helena Pahlm  
Quality Assurance Manager