

CERTIFICATE OF ANALYSIS

Product:	BabyBio Co-NTA 1 ml
Storage solution:	20% ethanol
Article number:	45 655 131, 45 655 132, 45 655 133, 45 655 134
Lot number:	154001
Resin lot number:	152906
Manufacturing date:	2020 JUL 08
Product release date:	2020 JUL 09
Expire date:	2021-09

	Measured values	Specifications
Colour:	Passed	Dark pink to purple
Metal ion capacity (Tested with Cu ²⁺)	22.5 µmol/ml	20 – 30 µmol/ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued Date: *20201012*

by


Helena Pahlm
Quality Assurance Manager