

## CERTIFICATE OF ANALYSIS

<b>Product:</b>	<b>BabyBio™ Dsalt 1 ml</b>
Storage solution:	<b>20% ethanol</b>
Article number:	<b>45 360 101, 45 360 102, 45 360 103, 45 360 104, 45 360 110</b>
Lot number:	<b>154053</b>
Resin lot number:	<b>CM01U0307H</b>
Manufacturing date:	<b>2020 AUG 28</b>
Product release date:	<b>2020 OCT 06</b>
Expire date:	<b>2023-MAY</b>

	<b>Measured values</b>	<b>Specifications</b>
Colour:	Passed	White to-off white
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2020 OCT 06**

by



Helena Pahlm

Quality Assurance Manager

