

CERTIFICATE OF ANALYSIS

Product:	BabyBio Cu-IDA 1 ml
Storage solution:	20% ethanol
Article number:	45 655 021, 45 655 022, 45 655 023, 45 655 024
Lot number:	154078
Resin lot number:	154017
Manufacturing date:	2020 SEP 18
Product release date:	2020 SEP 28
Expire date:	2025-05

	Measured values	Specifications
Capacity before charging with Cu ²⁺	60 µmol Cu ²⁺ /ml	50 – 70 µmol Cu ²⁺ /ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2020 SEP 29**

by



Helena Pahlm
Quality Assurance Manager