


## CERTIFICATE OF ANALYSIS

<b>Product:</b>	<b>BabyBio Cu-IDA 1 ml</b>
Storage solution:	<b>20% ethanol</b>
Article number:	<b>45 655 021, 45 655 022, 45 655 023, 45 655 024</b>
Lot number:	<b>154078</b> obsolete due to name change CC-22-10384
Resin lot number:	<b>154017</b> Philip Eriksson 2022-08-31
Manufacturing date:	<b>2020 SEP 18</b>
Product release date:	<b>2020 SEP 28</b>
Expire date:	<b>2025-05</b>

DocuSigned by:  
  
158C686C7E5E4A0...

	<b>Measured values</b>	<b>Specifications</b>
Capacity before charging with Cu <sup>2+</sup>	60 µmol Cu <sup>2+</sup> /ml	50 – 70 µmol Cu <sup>2+</sup> /ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2020 SEP 29**

by

  
Helena Pahlm  
Quality Assurance Manager