

## CERTIFICATE OF ANALYSIS

<b>Product:</b>	<b>BabyBio Cu-IDA 5 ml</b>
Storage solution:	<b>20% ethanol</b>
Article number:	<b>45 655 025, 45 655 026, 45 655 027, 45 655 028</b>
Lot number:	<b>154112</b>
Resin lot number:	<b>154017</b>
Manufacturing date:	<b>2020 SEP 22</b>
Product release date:	<b>2020 SEP 28</b>
Expire date:	<b>2025-07</b>

	<b>Measured values</b>	<b>Specifications</b>
Capacity before charging with Cu <sup>2+</sup>	60 µmol Cu <sup>2+</sup> /ml	50 – 70 µmol Cu <sup>2+</sup> /ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2020 SEP 29**

by



Helena Pahlm  
Quality Assurance Manager