

CERTIFICATE OF ANALYSIS

Product: BabyBio Cu-IDA 5 ml

Storage solution: 20% ethanol

Article number: 45 655 025, 45 655 026, 45 655 027, 45 655 028

Lot number: 154112 obsolete due to name change
CC-22-10384

Resin lot number: 154017

Manufacturing date: 2020 SEP 22 Philip Eriksson
2022-08-31

Product release date: 2020 SEP 28

Expire date: 2025-07

DocuSigned by:
Philip Eriksson
158C686C7E5E4A0...

	Measured values	Specifications
Capacity before charging with Cu ²⁺	60 µmol Cu ²⁺ /ml	50 – 70 µmol Cu ²⁺ /ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2020 SEP 29**

by


Helena Palm
Quality Assurance Manager