

CERTIFICATE OF ANALYSIS


Product:	BabyBio Cu-NTA 1 ml
Storage solution:	20% ethanol
Article number:	45 655 121, 45 655 122, 45 655 123, 45 655 124
Lot number:	154122
Resin lot number:	154022
Manufacturing date:	2020 OCT 01
Product release date:	2020 OCT 09
Expire date:	2025-05

	Measured values	Specifications
Colour:	Passed	Turquoise to light blue
Metal ion capacity (Tested with Cu ²⁺)	25 µmol/ml	20 – 30 µmol/ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2020 OCT 09**

by



Helena Pahlm
Quality Assurance Manager