


## CERTIFICATE OF ANALYSIS

<b>Product:</b>	<b>BabyBio Cu-NTA 1 ml</b>	
Storage solution:	<b>20% ethanol</b>	
Article number:	<b>45 655 121, 45 655 122, 45 655 123, 45 655 124</b>	
Lot number:	<b>154122</b>	obsolete due to name change CC-22-10384
Resin lot number:	<b>154022</b>	Philip Eriksson
Manufacturing date:	<b>2020 OCT 01</b>	2022-09-05
Product release date:	<b>2020 OCT 09</b>	
Expire date:	<b>2025-05</b>	

DocuSigned by:  
  
 158C686C7E5E4A0...

	Measured values	Specifications
Colour:	Passed	Turquoise to light blue
Metal ion capacity (Tested with Cu <sup>2+</sup> )	25 µmol/ml	20 – 30 µmol/ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2020 OCT 09**

by



Helena Pahlm  
 Quality Assurance Manager