

## CERTIFICATE OF ANALYSIS

<b>Product:</b>	<b>BabyBio™ A 1 ml</b>
Storage solution:	<b>20% ethanol</b>
Article number:	<b>45 605 101, 45 605 102, 45 605 103, 45 605 104</b>
Lot number:	<b>154129</b>
Resin lot number:	<b>153931</b>
Manufacturing date:	<b>2020 SEP 29</b>
Product release date:	<b>2020 OCT 01</b>
Expire date:	<b>2023-MAY</b>

	<b>Measured values</b>	<b>Specifications</b>
Dynamic binding capacity for IgG <sup>1</sup> .	50 mg/ml	> 40 mg/ml
Gap control: (4 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

1. Test done at 2.5 minutes residence time in 6.6x100 mm column.

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2020 OCT 01**

by



\_\_\_\_\_  
Helena Pahlm  
Quality Assurance Manager