


CERTIFICATE OF ANALYSIS


Product:	BabyBio Co-NTA 1 ml	
Storage solution:	20% ethanol	
Article number:	45 655 131, 45 655 132, 45 655 133, 45 655 134	
Lot number:	154137	obsolete due to name change CC-22-10384
Resin lot number:	154018	Philip Eriksson 2022-09-02
Manufacturing date:	2020 OCT 02	 158C686C7E5E4A0...
Product release date:	2020 OCT 12	
Expire date:	2025-05	

	Measured values	Specifications
Colour:	Passed	Dark pink to purple
Metal ion capacity (Tested with Cu ²⁺)	25 µmol/ml	20 – 30 µmol/ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2020 OCT 09**

by



Helena Pahlm

Quality Assurance Manager